



CS650 and CS655

Water Content Reflectometers



Revision: 08/2021 Copyright © 2011 – 2021 Campbell Scientific CSL I.D - 912





This equipment is guaranteed against defects in materials and workmanship. We will repair or replace products which prove to be defective during the guarantee period as detailed on your invoice, provided they are returned to us prepaid. The guarantee will not apply to:

- Equipment which has been modified or altered in any way without the written permission of Campbell Scientific
- Batteries
- Any product which has been subjected to misuse, neglect, acts of God or damage in transit.

Campbell Scientific will return guaranteed equipment by surface carrier prepaid. Campbell Scientific will not reimburse the claimant for costs incurred in removing and/or reinstalling equipment. This guarantee and the Company's obligation thereunder is in lieu of all other guarantees, expressed or implied, including those of suitability and fitness for a particular purpose. Campbell Scientific is not liable for consequential damage.

Please inform us before returning equipment and obtain a Repair Reference Number whether the repair is under guarantee or not. Please state the faults as clearly as possible, and if the product is out of the guarantee period it should be accompanied by a purchase order. Quotations for repairs can be given on request. It is the policy of Campbell Scientific to protect the health of its employees and provide a safe working environment, in support of this policy a "Declaration of Hazardous Material and Decontamination" form will be issued for completion.

When returning equipment, the Repair Reference Number must be clearly marked on the outside of the package. Complete the "Declaration of Hazardous Material and Decontamination" form and ensure a completed copy is returned with your goods. Please note your Repair may not be processed if you do not include a copy of this form and Campbell Scientific Ltd reserves the right to return goods at the customers' expense.

Note that goods sent air freight are subject to Customs clearance fees which Campbell Scientific will charge to customers. In many cases, these charges are greater than the cost of the repair.



Campbell Scientific Ltd, 80 Hathern Road, Shepshed, Loughborough, LE12 9GX, UK Tel: +44 (0) 1509 601141 Fax: +44 (0) 1509 270924

Email: support@campbellsci.co.uk www.campbellsci.co.uk Please note that this manual was originally produced by Campbell Scientific Inc. primarily for the North American market. Some spellings, weights and measures may reflect this origin.

Some useful conversion factors:

Area: 1 in^2 (square inch) = 64.	5 mm^2 Mass:	1 oz. (ounce) = 28.35 g 1 lb (pound weight) = 0.454 kg
Length: 1 in. (inch) = 25.4 m 1 ft (foot) = 304.8 m 1 yard = 0.914 m	m Pressure:	$1 \text{ psi} (\text{lb/in}^2) = 68.95 \text{ mb}$
1 mile = 1.609 km	Volume:	1 UK pint = 568.3 ml 1 UK gallon = 4.546 litres 1 US gallon = 3.785 litres

In addition, while most of the information in the manual is correct for all countries, certain information is specific to the North American market and so may not be applicable to European users.

Differences include the U.S standard external power supply details where some information (for example the AC transformer input voltage) will not be applicable for British/European use. *Please note, however, that when a power supply adapter is ordered it will be suitable for use in your country.*

Reference to some radio transmitters, digital cell phones and aerials may also not be applicable according to your locality.

Some brackets, shields and enclosure options, including wiring, are not sold as standard items in the European market; in some cases alternatives are offered. Details of the alternatives will be covered in separate manuals.

Part numbers prefixed with a "#" symbol are special order parts for use with non-EU variants or for special installations. Please quote the full part number with the # when ordering.

Recycling information



At the end of this product's life it should not be put in commercial or domestic refuse but sent for recycling. Any batteries contained within the product or used during the products life should be removed from the product and also be sent to an appropriate recycling facility.

Campbell Scientific Ltd can advise on the recycling of the equipment and in some cases arrange collection and the correct disposal of it, although charges may apply for some items or territories.

For further advice or support, please contact Campbell Scientific Ltd, or your local agent.



Campbell Scientific Ltd, 80 Hathern Road, Shepshed, Loughborough, LE12 9GX, UK Tel: +44 (0) 1509 601141 Fax: +44 (0) 1509 270924 Email: support@campbellsci.co.uk www.campbellsci.co.uk

Safety

DANGER — MANY HAZARDS ARE ASSOCIATED WITH INSTALLING, USING, MAINTAINING, AND WORKING ON OR AROUND **TRIPODS, TOWERS, AND ANY ATTACHMENTS TO TRIPODS AND TOWERS SUCH AS SENSORS, CROSSARMS, ENCLOSURES, ANTENNAS, ETC**. FAILURE TO PROPERLY AND COMPLETELY ASSEMBLE, INSTALL, OPERATE, USE, AND MAINTAIN TRIPODS, TOWERS, AND ATTACHMENTS, AND FAILURE TO HEED WARNINGS, INCREASES THE RISK OF DEATH, ACCIDENT, SERIOUS INJURY, PROPERTY DAMAGE, AND PRODUCT FAILURE. TAKE ALL REASONABLE PRECAUTIONS TO AVOID THESE HAZARDS. CHECK WITH YOUR ORGANIZATION'S SAFETY COORDINATOR (OR POLICY) FOR PROCEDURES AND REQUIRED PROTECTIVE EQUIPMENT PRIOR TO PERFORMING ANY WORK.

Use tripods, towers, and attachments to tripods and towers only for purposes for which they are designed. Do not exceed design limits. Be familiar and comply with all instructions provided in product manuals. Manuals are available at www.campbellsci.eu or by telephoning +44(0) 1509 828 888 (UK). You are responsible for conformance with governing codes and regulations, including safety regulations, and the integrity and location of structures or land to which towers, tripods, and any attachments are attached. Installation sites should be evaluated and approved by a qualified engineer. If questions or concerns arise regarding installation, use, or maintenance of tripods, towers, attachments, or electrical connections, consult with a licensed and qualified engineer or electrician.

General

- Prior to performing site or installation work, obtain required approvals and permits. Comply with all governing structure-height regulations, such as those of the FAA in the USA.
- Use only qualified personnel for installation, use, and maintenance of tripods and towers, and any attachments to tripods and towers. The use of licensed and qualified contractors is highly recommended.
- Read all applicable instructions carefully and understand procedures thoroughly before beginning work.
- Wear a hardhat and eye protection, and take other appropriate safety precautions while working on or around tripods and towers.
- **Do not climb** tripods or towers at any time, and prohibit climbing by other persons. Take reasonable precautions to secure tripod and tower sites from trespassers.
- Use only manufacturer recommended parts, materials, and tools.

Utility and Electrical

- You can be killed or sustain serious bodily injury if the tripod, tower, or attachments you are installing, constructing, using, or maintaining, or a tool, stake, or anchor, come in contact with overhead or underground utility lines.
- Maintain a distance of at least one-and-one-half times structure height, or 20 feet, or the distance required by applicable law, whichever is greater, between overhead utility lines and the structure (tripod, tower, attachments, or tools).
- Prior to performing site or installation work, inform all utility companies and have all underground utilities marked.
- Comply with all electrical codes. Electrical equipment and related grounding devices should be installed by a licensed and qualified electrician.

Elevated Work and Weather

- Exercise extreme caution when performing elevated work.
- Use appropriate equipment and safety practices.
- During installation and maintenance, keep tower and tripod sites clear of un-trained or non-essential personnel. Take precautions to prevent elevated tools and objects from dropping.
- Do not perform any work in inclement weather, including wind, rain, snow, lightning, etc.

Maintenance

- Periodically (at least yearly) check for wear and damage, including corrosion, stress cracks, frayed cables, loose cable clamps, cable tightness, etc. and take necessary corrective actions.
- Periodically (at least yearly) check electrical ground connections.

WHILE EVERY ATTEMPT IS MADE TO EMBODY THE HIGHEST DEGREE OF SAFETY IN ALL CAMPBELL SCIENTIFIC PRODUCTS, THE CUSTOMER ASSUMES ALL RISK FROM ANY INJURY RESULTING FROM IMPROPER INSTALLATION, USE, OR MAINTENANCE OF TRIPODS, TOWERS, OR ATTACHMENTS TO TRIPODS AND TOWERS SUCH AS SENSORS, CROSSARMS, ENCLOSURES, ANTENNAS, ETC.

Table of contents

1. Introduction	1
2. Precautions	1
3. Initial inspection	1
4. QuickStart	2
5. Overview	4
6. Specifications	6
6.1 Electrical specifications	7 8
7. Installation	10
7.1 Orientation and placement 7.2 Proper insertion	10 10
7.3 Data logger wiring 7.4 Programming	11 12
8. Operation	13
8. Operation 8.1 A200 and Device Configuration Utility	13 13
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 	13 13 13
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1.1 Driver installation 	13 13 13 13
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1.1 Driver installation 8.1.1.2 RS-232 wiring 	13 13 13 13 13
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1.1 Driver installation 8.1.1.2 RS-232 wiring 8.1.1.3 Powering the sensor 	13 13 13 13 13 14
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1.1 Driver installation 8.1.1.2 RS-232 wiring 8.1.1.3 Powering the sensor 8.1.1.4 Determining which COM port the A200 has been assigned 	13 13 13 13 13 14 14
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1 Driver installation 8.1.1.2 RS-232 wiring 8.1.1.3 Powering the sensor 8.1.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 	13 13 13 13 13 14 14 14
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1 Driver installation 8.1.1.2 RS-232 wiring 8.1.1.3 Powering the sensor 8.1.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2.1 Settings Editor tab 	13 13 13 13 13 14 14 14 15
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1.1 Driver installation 8.1.2 RS-232 wiring 8.1.3 Powering the sensor 8.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2.1 Settings Editor tab 8.1.2.2 Send OS tab 	13 13 13 13 14 14 14 15 18
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1 Driver installation 8.1.2 RS-232 wiring 8.1.3 Powering the sensor 8.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2.1 Settings Editor tab 8.1.2.2 Send OS tab 8.2 SDI-12 measurements 	13 13 13 13 13 14 14 14 14 15 18 19
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1.1 Driver installation 8.1.1.2 RS-232 wiring 8.1.1.3 Powering the sensor 8.1.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2.1 Settings Editor tab 8.1.2.2 Send OS tab 8.2 SDI-12 measurements 8.2.1 M3! and M4! commands 	13 13 13 13 13 14 14 14 14 15 18 19 20
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1 Driver installation 8.1.1.2 RS-232 wiring 8.1.1.3 Powering the sensor 8.1.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2.1 Settings Editor tab 8.1.2.2 Send OS tab 8.2 SDI-12 measurements 8.2.1 M3! and M4! commands 8.2.2 Use of multiplexers 	13 13 13 13 14 14 14 15 18 19 20 22
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1 Driver installation 8.1.2 RS-232 wiring 8.1.3 Powering the sensor 8.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2 Settings Editor tab 8.1.2.2 Send OS tab 8.2 SDI-12 measurements 8.2.1 M3! and M4! commands 8.2.2 Use of multiplexers 8.3 Water content reflectometer method for measuring volumetric water content 	13 13 13 13 14 14 14 15 18 19 20 22 22
 8. Operation 8.1 A200 and Device Configuration Utility 8.1.1 Using the A200 8.1.1 Driver installation 8.1.2 RS-232 wiring 8.1.3 Powering the sensor 8.1.4 Determining which COM port the A200 has been assigned 8.1.2 Device Configuration Utility 8.1.2.1 Settings Editor tab 8.1.2.2 Send OS tab 8.2 SDI-12 measurements 8.2.1 M3! and M4! commands 8.2.2 Use of multiplexers 8.3 Water content reflectometer method for measuring volumetric water content 8.3.1 Description of measurement method 	13 1313131414141415181920222222

8.3.3 Electrical conductivity	23
8.3.3.1 Soil electrical conductivity	23
8.3.3.2 Temperature correction of soil electrical conductivity	24
8.3.4 Error sources in water content reflectometer measurement	24
8.3.4.1 Sensor-to-sensor variability error	24
8.3.4.2 Insertion error	
8.3.5 Temperature dependence and correction	
8.3.5.1 Accurate soil temperature measurement	26
8.4 Water content reflectometer user-calibration	
8.4.1 Need for soil specific calibration equation	
8.4.2 User-derived calibration equation	26
8.4.3 Collecting laboratory data for calibration	27
8.4.4 Collecting field data for calibration	
8.4.5 Calculations	
9. Maintenance and troubleshooting	
10. References	
Appendix A. Importing Short Cut code into CRBasic Editor	
Appendix B. Example programs	
Appendix B. Example programs B.1 CR1000X with one CS650 sensor	36 36
Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port	36
Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content 	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support 	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics 	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) 	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) D.1.2 Send identification command (a!) 	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) D.1.2 Send identification command (aV!) 	36 37 38 40 41 41 42 42 42 43
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) D.1.2 Send identification command (a!) D.1.3 Start verification command (aV!) D.1.4 Address query command (?!) 	
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) D.1.2 Send identification command (a!) D.1.3 Start verification command (aV!) D.1.4 Address query command (?!) D.1.5 Change address command (aAb!) 	36 37 38 40 41 41 42 42 42 43 43 43
 Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) D.1.2 Send identification command (a!) D.1.3 Start verification command (aV!) D.1.4 Address query command (?!) D.1.5 Change address command (aAb!) D.1.6 Start measurement commands (aM!) 	36 37 38 40 41 41 42 42 42 42 43 43 43 43
Appendix B. Example programsB.1 CR1000X with one CS650 sensorB.2 CR1000X with two CS650 sensors on same control portB.3 CR1000X with 12 CS650 sensors on multiplexerAppendix C. Discussion of soil water contentAppendix D. SDI-12 sensor supportD.1 SDI-12 command basicsD.1.1 Acknowledge active command (a!)D.1.2 Send identification command (a!!)D.1.3 Start verification command (aV!)D.1.4 Address query command (?!)D.1.5 Change address command (aAb!)D.1.6 Start measurement commands (aM!)D.1.7 Stopping a measurement command	36 37 38 40 41 41 42 42 42 43 43 43 43 43 43
Appendix B. Example programsB.1 CR1000X with one CS650 sensorB.2 CR1000X with two CS650 sensors on same control portB.3 CR1000X with 12 CS650 sensors on multiplexerAppendix C. Discussion of soil water contentAppendix D. SDI-12 sensor supportD.1 SDI-12 command basicsD.1.1 Acknowledge active command (a!)D.1.2 Send identification command (a!!)D.1.3 Start verification command (aV!)D.1.4 Address query command (?!)D.1.5 Change address command (aAb!)D.1.6 Start measurement commands (aM!)D.1.7 Stopping a measurement commandD.1.8 Send data command (aD0! aD9!)	36 37 38 40 41 41 42 42 42 43 43 43 43 43 43 43
Appendix B. Example programs B.1 CR1000X with one CS650 sensor B.2 CR1000X with two CS650 sensors on same control port B.3 CR1000X with 12 CS650 sensors on multiplexer Appendix C. Discussion of soil water content Appendix D. SDI-12 sensor support D.1 SDI-12 command basics D.1.1 Acknowledge active command (a!) D.1.2 Send identification command (a!) D.1.3 Start verification command (aV!) D.1.4 Address query command (?!) D.1.5 Change address command (aAb!) D.1.6 Start measurement commands (aM!) D.1.7 Stopping a measurement command D.1.8 Send data command (aD0! aD9!) D.2 SDI-12 transparent mode	36 37 38 40 41 41 42 42 42 42 43 43 43 43 43 43 43 43 43

D.3 References	.7
----------------	----

1. Introduction

The CS650 and CS655 are multiparameter smart sensors that use innovative techniques to monitor soil volumetric water content, bulk electrical conductivity, and temperature. They output an SDI-12 signal that many Campbell Scientific data loggers can measure.

The CS650 has 30 cm length rods, whereas the CS655 has 12 cm length rods. This manual uses CS650 to reference model numbers CS650 and CS655. Unless specifically stated otherwise, information in the manual applies equally to both models.

NOTE:

This manual provides information only for CRBasic data loggers. For retired Edlog data logger support, see an older manual at www.campbellsci.com/old-manuals

2. Precautions

- READ AND UNDERSTAND the Safety section at the front of this manual.
- Although the CS650 is rugged, it should be handled as a precision scientific instrument.
- External radio frequency (RF) sources can affect the sensor operation. Therefore, the sensor should be located away from significant sources of RF such as ac power lines and motors.

3. Initial inspection

- Upon receipt of the CS650, inspect the packaging and contents for damage. File damage claims with the shipping company.
- The model number and cable length are printed on a label at the connection end of the cable. Check this information against the shipping documents to ensure the expected product and cable length were received.

4. QuickStart

A video that describes data logger programming using *Short Cut* is available at: www.campbellsci.eu/videos/cr1000x-data logger-getting-started-program-part-3. Short *Cut* is an easy way to program your data logger to measure the sensor and assign data logger wiring terminals. *Short Cut* is available as a download on www.campbellsci.eu. It is included in installations of *LoggerNet*, *RTDAQ*, and *PC400*.

The following procedure also describes programming with Short Cut.

- 1. Open Short Cut and click Create New Program.
- 2. Double-click the data logger model.
- 3. In the Available Sensors and Devices box, type CS650. You can also locate the sensor in the Sensors > Meteorological > Soil Moisture > CS650/CS655 Water Content Reflectometer folder. The four different options monitor different parameters. For this tutorial, double-click to select CS650/CS655 Water Content Reflectometer (VWC, EC, T, P, PA, and VR). The soil temperature defaults to degree C, and the sensor is measured hourly. These are changed by clicking the Temperature or Measure sensor box and selecting an option. Enter the correct SDI-12 Address for the CS650 or CS655. After entering the Properties, click on the Wiring tab to see how the sensor is to be wired to the data logger.



Properties Wiring		
	CS650/655 (VWC, EC, T, P, PA, and VR)	CR1000X Series
	Red	12V
	Green	C1
	Orange	G
	Black	G
	Clear	G
	Click a CR1000X Series terminal name to change a	wire's location.

- 4. Repeat step three for other sensors you want to measure.
- 5. In **Output Setup**, type the scan rate. If you chose to measure the CS650 hourly rather than every scan, this scan interval must be evenly divisible into an hour. Type a meaningful table name and type the **Data Output Storage Interval**.

Progress	How often should the CR1000X Series measure its	G
1. New/Open 2. Datalogger	sensor(s)?	
3. Sensors		
4. Output Setup	Data is processed by the datalogger and then stored in an output table. Two tables	
5. Adv. Outputs	are defined by default; up to 10 tables can	(
6. Output Select	be added.	
7. Finish	<u>1</u> Hourly <u>2</u> Daily	
Winin a	Table Name	
Wiring Diagram	Daily 😑 Delete Table	0
Wiring Text	Data Output Storage Interval	
	Makes 96 measurements per output	-
	measurement interval of 15 Minutes.	
	Advanced Outputs (all tables)	
	Specify how often measurements are to be made and how often outputs are to be stored. Note	
	that multiple output intervals can be specified, one for each output table. By default, an output table is set up to send data to memory based on time. Select the Advanced Output option to se	nd
	data to memory based on one or more of the following conditions: time, the state of a flag, or the	e
		_
	Previous Next Finish Help	

6. S	elect the m	neasurement ar	nd its ass	ociated (output op	tions.
------	-------------	----------------	------------	-----------	-----------	--------

1. New/Open	Selected Measurem for Output	ents Available		Selected Me	easuremen	ts for Outp	ut	
2. Datalogger	Sensor	Measurement	Average	1 Table1	2 Table2			
3 Sensors	▲ CR1000X		ETo	Sensor	easuremer	Processing	utput Labe	Units
4 Output Setup	▲ Default	BattV	Maximum	CS650/655	VWC	Average	VWC AVG	m^3/m^3
F. Adv. Outputs	- Inc	PTemp_C	Minimum	CS650/655	FC	Average	EC AVG	dS/m
6 Output Select	▲ CS650/655 (V	VWC	Minindin	CS650/655	т	Average	T AVG	Deg C
6. Output Select		EC	Sample	C3030/032	- -	Average	T_AVG	Deg C
7. Finish		т	StdDev	CS650/655	P	Average	P_AVG	unitiess
ANN 10	-	P	Total	CS650/655	PA	Average	PA_AVG	nSec
Viring		PA	WindVector	CS650/655	VR	Average	VR_AVG	unitless
Wiring Text								
				🖉 Edit	😦 Rem	ove		

- 7. Click **Finish** and save the program. Send the program to the data logger if the data logger is connected to the computer.
- 8. If the sensor is connected to the data logger, check the output of the sensor in the data display in *LoggerNet*, *RTDAQ*, or *PC400* to make sure it is making reasonable measurements.

5. Overview

The CS650 measures volumetric water content, electrical conductivity, dielectric permittivity, and temperature of soils or other porous media. These values are reported through SDI-12 communications.



FIGURE 5-1. CS650 (left) and CS655 Water Content Reflectometers

Volumetric water content (VWC) is derived from the sensor sensitivity to the dielectric permittivity of the medium surrounding the sensor stainless-steel rods. The CS650 is configured as a water content reflectometer, with the two parallel rods forming an open-ended transmission line. A differential oscillator circuit is connected to the rods, with an oscillator state change triggered by the return of a reflected signal from one of the rods. The two-way travel time of the electromagnetic waves that are induced by the oscillator on the rod varies with changing dielectric permittivity. Water is the main contributor to the bulk dielectric permittivity of the soil or porous media, so the travel time of the reflected wave increases with increasing water content and decreases with decreasing water content, hence the name water content reflectometer. The average travel time of the reflected wave multiplied by a scaling factor of 128 is called the period average. Period average is reported in microseconds and is considered to be the raw output of a water content reflectometer.

Electrical conductivity is determined by exciting the rods with a known non-polarizing waveform and measuring the signal attenuation. Signal attenuation is reported as a dimensionless voltage ratio, which is the ratio of the excitation voltage to the measured voltage along the sensor rods when they are excited at a fixed 100 kHz frequency. Voltage ratio ranges from 1 in nonconductive media to about 17 in highly conductive media. Values greater than 17 are highly unstable and indicate that the soil conditions are outside of the specified operating range of the sensor.

Temperature is measured with a thermistor in contact with one of the rods.

It is well known that transmission line oscillators used for water content measurements suffer from unwanted increases in oscillation period as increasing electrical conductivity causes

transmission line signal attenuation. The CS650 handles this problem by making an electrical conductivity measurement then correcting the oscillator period accordingly. On-board processing within the sensor head calculates electrical conductivity from the signal attenuation measurement and combines the result with the oscillation period measurement to calculate the dielectric permittivity of the media and finally applies the Topp equation (Topp et al. 1980) to estimate volumetric water content.

Sensor electronics are encapsulated in the rugged epoxy sensor head.

A five-conductor cable including the drain or shield wire provides power and ground as well as communications with the CS650. The CS650 is intended to communicate with SDI-12 recorders, including Campbell Scientific data loggers. Alternatively, use the orange wire to transmit and receive RS-232. The A200 USB-to-Serial Module allows RS-232 communications between a computer and the CS650 by means of Campbell Scientific's *Device Configuration Utility* software.

Features:

- Measurement corrected for effects of soil texture and electrical conductivity
- Estimates soil-water content for a wide range of mineral soils
- Versatile sensor—measures dielectric permittivity, bulk electrical conductivity (EC), and soil temperature
- Compatible with Campbell Scientific CRBasic data loggers: CR6 series, CR3000, CR1000X series, CR800 series, CR300 series, and CR1000

6. Specifications

Table 6-1 (p. 6) compares the size of the CS650 and CS655 reflectometers.

Table 6-1: Size specifications			
	CS650	CS655	
Rods	300 mm long 3.2 mm diameter 32 mm spacing	120 mm long 3.2 mm diameter 32 mm spacing	
Sensor head	L 85 mm W 63 mm D 18 mm	L 85 mm W 63 mm D 18 mm	

Table 6-1: Size specifications			
	CS650 CS655		
Sensor weight	280 g	240 g	
Cable weight	35 g m ⁻¹	35 g m ⁻¹	

Ingress protection rating:

IP68

6.1 Electrical specifications

Sensor output:	SDI-12 Serial RS-232	
Warmup time:	3 s	
Measurement time:	3 ms to measure 600 ms to complete SDI-12 command	
Power supply requirements:	6 to 18 VDC Must be able to supply 45 mA @ 12 VDC	
Maximum cable length:	610 m (2000 ft) combined length for up to 25 sensors connected to the same data logger control or U terminal ¹	
Electromagnetic compatibility:	View EU Declaration of Conformity documentation at www.campbellsci.eu/cs650	
	External RF sources can affect CS650 measurements. CS650 circuitry should be located away from radio transmitter aerials and cables, or measurements discarded during RF transmissions.	
Active current drain (3 ms):	45 mA typical @ 12 VDC (80 mA @ 6 VDC, 35 mA @ 18 VDC)	
Quiescent current drain:	135 mA @ 12 VDC	
Average current drain:	I = 0.09n + [3.5 + 0.024(n-1)]n/s I = average current in milliamps n = number of CS650s s = number of seconds between measurements (see FIGURE 6-1 (p. 8))	

¹Campbell Scientific recommends using separate terminals when possible.



FIGURE 6-1. CS650 and CS655 average current drain

FIGURE 6-1 (p. 8) shows average current drain for different measurement rates and quantities of CS650 sensors. If the time between measurements is five minutes or longer, average current drain is approximately 0.15 milliamps per sensor.

6.2 Operational specifications

Table 6-2 (p. 8) provides the operational specifications.

Table 6-2: Relative dielectric permittivity specifications				
	CS650	CS655		
Relative dielectric permittivity				
Range	1 to 81			
Accuracy ⁺				
1 to 40:	\pm (2% of reading + 0.6) for solution electrical conductivity (EC) ≤3 dS/m	\pm (3% of reading + 0.8) for solution EC \leq 8 dS/m		
40 to 80:	\pm 1.4 for solution EC \leq 3 dS/m	± 2 for solution EC ≤ 2.8 dS/m		
Precision‡	<0.02			

Table 6-2: Relative diele	ctric permittivity specifications		
	CS650	CS655	
Volumetric water conter	nt		
Range	0% to	100%	
Accuracy†	±1% (with soil specific calibration), ±3% (typical with factory VWC model) where solution EC < 3 dS/m	±1% (with soil specific calibration), ±3% (typical with factory VWC model) where solution EC < 10 dS/m	
Precision‡	<0.05%		
Electrical conductivity			
Range solution EC	0 to 3 dS/m	0 to 8 dS/m	
Range bulk EC	0 to 3 dS/m	0 to 8 dS/m	
Accuracy ⁺	±(5% of reading + 0.05 dS/m)		
Precision‡	0.5% of bulk electrical conductivity (BEC)		
Temperature			
Range	–50 tc	o 70 °C	
Resolution	0.00)1 °C	
Accuracy†	±0.1 °C (for typica [0 to 40 °C] when sensc ±0.5 °C for full te	l soil temperatures or body is buried in soil), mperature range	
Precision [‡]	±0.02 °C		
Sensing volume*	7800 cm ³	3600 cm ³	
*Sensing volume approximat †Accuracy specifications are b ranging from 1 to 81 and solu	ely 7.5 cm radius around each sensor rod a based on laboratory measurements in a seri tion electrical conductivities ranging from 0	nd 4.5 cm beyond the end of the rods es of solutions with dielectric permittivities to 3 dS/m.	

‡Precision describes the repeatability of a measurement. It is determined for the CS650 by taking repeated measurements in the same material. The precision of the CS650 is better than 0.05 % volumetric water content and 0.01 dS/m electrical conductivity.

7. Installation

If you are programming your data logger with *Short Cut*, skip Data logger wiring (p. 11) and Programming (p. 12). *Short Cut* does this work for you. See QuickStart (p. 2) for a *Short Cut* tutorial.

7.1 Orientation and placement

The CS650 measures the bulk dielectric permittivity, average volumetric water content, and bulk EC along the length of the rods, which is 30 cm for the CS650 and 12 cm for the CS655. The sensor rods can insert vertically into the soil surface or buried at any orientation to the surface. Install the sensor horizontal to the surface to detect the passing of wetting fronts or other vertical water fluxes.

The sensitive volume depends on the surrounding media. In soil, the sensitive volume extends approximately 7.5 cm (3 in) from the rods along their length and 4.5 cm (1.8 in) beyond the end of the rods. Consequently, if the sensor is buried horizontally closer than 7.5 cm from the soil surface, it includes air above the surface in its measurements and underestimates soil water content.

The thermistor used to measure temperature is in contact with one of the stainless steel rods at the base of the epoxy sensor body. Because of the low thermal conductivity of stainless steel, the thermistor does not measure the average temperature along the rod, but instead provides a point measurement of the temperature within the epoxy. For a valid soil temperature reading, the sensor body must be in thermal equilibrium with the soil. Also bury the sensor head in the soil so that it is insulated from diurnal temperature fluctuations.

7.2 Proper insertion

The method used for sensor installation affects the accuracy of the measurement. Insert the sensor rods as close to parallel as possible. The sensor is more sensitive to permittivity close to the rods. Air voids around the rods result in reduced measurement accuracy. Most soils recover from disturbances caused by installation.

The CS650G Rod Insertion Guide Tool with Pilot Rod helps maintain the proper spacing and parallel orientation of the rods during sensor insertion. It also helps the insertion of the sensor in high density or rocky soils.



FIGURE 7-1. CS650G Rod Insertion Guide Tool with Pilot Rod

7.3 Data logger wiring

Campbell Scientific data loggers typically use SDI-12 to measure the sensor because RS-232 communications require more terminals per CS650 and RS 232 programming is more complicated than SDI-12 programming.

Table 7-1 (p. 11) shows the SDI-12 wiring for the CS650 water content reflectometer. SDI-12 data is transmitted to a CRBasic data logger odd numbered control or U terminal. Wiring information for RS-232 communications is provided in RS-232 wiring (p. 13).

Table 7-1: Wire colour, function, and data logger connection for SDI-12		
Colour	Function	Data logger connection
Green	SDI-12 data	C , SDI-12, or U configured for SDI-12 ¹
Red	SDI-12 power	12V
Black	SDI-12 reference	G
Clear	Shield	G
Orange	Not used	G
¹ U terminals are automatically configured by the measurement instruction.		

NOTE:

The orange wire is only used for RS-232 communications and should be grounded when using SDI-12.

For the CR6 and CR1000X data loggers, triggering conflicts may occur when a companion terminal is used for a triggering instruction such as TimerInput(), PulseCount(), or WaitDigTrig(). For example, if the CS650 is connected to C3 on a CR1000X, C4 cannot be used in the TimerInput(), PulseCount(), or WaitDigTrig() instructions.

7.4 Programming

Short Cut is the best source for up-to-date programming code for Campbell Scientific data loggers. If your data acquisition requirements are simple, you can probably create and maintain a data logger program exclusively with *Short Cut*. If your data acquisition needs are more complex, the files that *Short Cut* creates are a great source for programming code to start a new program or add to an existing custom program.

NOTE:

Short Cut cannot edit programs after they are imported and edited in CRBasic Editor.

A *Short Cut* tutorial is available in QuickStart (p. 2). If you wish to import *Short Cut* code into *CRBasic Editor* to create or add to a customized program, follow the procedure in Importing Short Cut code into CRBasic Editor (p. 35). Programming basics for CRBasic data loggers are provided in this section. Complete program examples for select CRBasic data loggers can be found in Example programs (p. 36).

Use the **SDI12Recorder()** instruction to measure the sensor. This instruction sends a request to the sensor to make a measurement then retrieves the measurement values from the sensor. See SDI-12 measurements (p. 19) for more information.

For most CRBasic data loggers, the **SDI12Recorder()** instruction has the following syntax:

SDI12Recorder(Destination, SDIPort, SDIAddress, "SDICommand", Multiplier, Offset, FillNAN, WaitonTimeout)

The **Destination** parameter must be an array. The required number of values in the array depends on the command (Table 8-3 (p. 19)).

FillNAN and **WaitonTimeout** are optional parameters (refer to CRBasic Help for more information).

8. Operation

8.1 A200 and *Device Configuration Utility*

The A200 Sensor-to-Computer Interface allows communications between a CS650 and a computer to change sensor settings through *Device Configuration Utility* software.

8.1.1 Using the A200

8.1.1.1 Driver installation

If the A200 has not been previously plugged into your computer and your computer operating system is not Windows 7, the A200 driver needs to be installed.

NOTE:

Drivers should be installed before plugging the A200 into the computer.

To download the A200 drivers, go to www.campbellsci.eu/downloads/a200-sensor-interface-usb-driver and click **Download Now**.

8.1.1.2 RS-232 wiring

One end of the A200 has a terminal block while the other end has a type B female USB port. The terminal block provides **12V**, **G**, **Tx**, and **Rx** terminals for connecting the sensor (see FIGURE 8-1 (p. 13) and Table 8-1 (p. 14)).



FIGURE 8-1. A200 Sensor-to-Computer Interface

A data cable that ships with the A200 has a USB type-A male connector for attaching to a computer USB port, and a type B male connector for attaching to the A200 USB port.

Table 8-1: CS650 wiring code for RS-232 and A200		
Colour	Function	A200 terminal
Orange	RxD	Rx
Green	TxD	Тх
Red	Power	+12 VDC
Black	Reference	G
Clear	Shield	G

8.1.1.3 Powering the sensor

The A200 provides power to the sensor when it is connected to a computer USB port. An internal dc/dc converter boosts the 5 VDC supply from the USB connection to a 12 VDC output that is required to power the sensor.

8.1.1.4 Determining which COM port the A200 has been assigned

When the A200 driver is installed, the A200 is assigned a COM port number. This COM port number is needed when using *Device Configuration Utility*. Often, the assigned COM port is the next port number that is free. However, if other devices were previously installed (some of which may not be plugged in), the A200 may be assigned a higher COM port number.

To check which COM port has been assigned to the A200, monitor the appearance of a new COM port in the list of COM ports offered in your software package such as *LoggerNet* before and after the installation, or look in the Windows Device Manager list under the ports section (access via the control panel).

8.1.2 Device Configuration Utility

The *Device Configuration Utility* is available as a download on www.campbellsci.eu.downloads. It is included in installations of *LoggerNet*, *RTDAQ*, or *PC400*.

Connect the CS650 to the A200 as shown in Table 8-1 (p. 14). Connect the computer to the A200 USB port with the supplied USB cable.

Launch *Device Configuration Utility* and search for CS650 Series from the Device Type list on the left. Select 9600 from the Baud Rate list.



Browse for and select the Communications Port used for your device. (see Determining which COM port the A200 has been assigned (p. 14)).

Select Ok then Connect to begin communications with the CS650.

8.1.2.1 Settings Editor tab

The **Settings Editor** tab shows settings stored in the CS650 operating system. Settings that may be modified include **User Name**, **SDI-12 Address**, and **RS-232 Baud Rate**. Attempts to change any of the other settings results in a "Commit failed. Unrecognized error condition" error message. *Device Configuration Utility* polls the CS650 every two seconds while connected and the results are displayed under **Real-Time Measurements** (Table 8-2 (p. 18)). This is useful for verifying sensor performance.

DS Version	
CS65x.Std.00.22	
DS Date	
March 19,2010 3:30pm	
Model Name	
CS650 V	
Serial Number	
0	
Jser Name	
No User Name	
DI 12 Address	
0 +	
RS-232 Baud Rate	
9600 🗸	
Real-Time Measurements	
Period(uSec) = 1.459	^
Maltana Dalia 1	
voltage Ratio = 1	
Dielectric Permitivity = 6.698	~
(a Multipliar	
30	
(a Offset	
2.975228	
EC Multiplier	
-2.09345	
EC Offret	
Le onset	

Default communications settings are 9600 baud, no parity, 1 stop bit, 8 data bits, and no error checking. After any changes to CS650 settings, select **Apply** to write the changes to the CS650 operating system. A configuration summary is then shown. The summary may be printed or saved electronically for future reference.

Configuration of configured on: Monday, Febr	of CS650 Series	
Setting Name	Setting Value	
	CS65x.Std.00.22	
	March 19,2010 3:30pm	
Model_Name	CS650	
Sensor SN	4,294,967,295	
	No User Name	
SDI-12_Address	0	
Baud_Rate	9600	
	Measurement Field Name	Measurements
	Period(uSec) =	1.459
	Voltage Ratio =	1
Real-Time Measurements	Dielectric Permitivity =	6.698
	Volumetric Water Content =	0.1192
	Electrical Conductivity(dS/m) =	0.0204
	Temperature(°C) =	26.16
Ka_Mult	30	

Table 8-2: Real-Time Measurements		
Measurement field name	Meaning	
VWC	Volumetric water content	
EC (dS/m)	Bulk electrical conductivity	
TS (°C)	Soil temperature	
Ка	Bulk dielectric permittivity	
PA (μS)	Period average	
VR	Voltage ratio	

8.1.2.2 Send OS tab

The **Send OS** tab is used to update the operating system in the CS650. The operating system is available at www.campbellsci.eu/downloads. The file to send has a filename extension of .a43, such as CS65X.Std.02.a43. Sending a new operating system does not affect any of the user-modified settings or sensor specific multiplier and offset settings.

To download a new operating system, follow the **Operating System Download Procedure** listed on the **Send OS** tab.



8.2 SDI-12 measurements

The CS650 responds to SDI-12 commands M!, M1!, M2!, M3!, M4!, ?!, and I!. Table 8-3 (p. 19) shows the values returned for each of these commands.

Establish SDI-12 communications using the **SDI12Recorder()** CRBasic instruction. See Data logger wiring (p. 11) for SDI-12 wiring details.

NOTE:

This section briefly describes using SDI-12 commands. Additional SDI-12 information is available at SDI-12 sensor support (p. 41), www.sdi-12.org \Box , or www.youtube.com/user/CampbellScientific \Box .

Table 8-3: CS650 SDI-12 comma	nds	
SDI-12 command ("a" is the sensor address)	Values returned	Units
	1. Volumetric Water Content, q	m ³ /m ³
aM!	2. Electrical Conductivity, s	dS/m
	3. Temperature	°C
	1. Permittivity, e	
aM1!	2. Electrical Conductivity, s	dS/m
	3. Temperature	°C
	1. Period, t	μs
aM2 !	2. Voltage Ratio, a	
	3. Temperature	°C
	1. Volumetric Water Content, q	m ³ /m ³
	2. Electrical Conductivity, s	dS/m
aM3!	3. Temperature	
	4. Permittivity, e	°C
	5. Period, t	μs
	6. Voltage Ratio, a	

Table 8-3: CS650 SDI-12 comma	inds	
SDI-12 command ("a" is the sensor address)	Values returned	Units
	1. Volumetric Water Content, q	m ³ /m ³
	2. Electrical Conductivity, s	dS/m
эM4 I	3. Temperature	°C
a194 !	4. Permittivity, e	
	5. Period, t	μs
	6. Voltage Ratio, a	
aM5! aM9!	No Values Returned	
?!	Returns the SDI-12 Address	
aI!	CampbellSci, Operating System Version, Product Serial Number	

Campbell Scientific recommends using separate terminals when possible. However, multiple CS650 sensors can connect to the same data logger control or **U** terminal. Each must have a unique SDI-12 address. Valid addresses are 0 through 9, A through Z, and a through z. The CS650 ships with a default SDI-12 address of 0 unless otherwise specified at the time of ordering. Change the SDI-12 address through *Device Configuration Utility* (see A200 and Device Configuration Utility (p. 13)) or with a terminal emulator in SDI-12 transparent mode (see SDI-12 sensor support (p. 41)).

8.2.1 M3! and M4! commands

Typically, the M4! command is used to report volumetric water content, electrical conductivity, temperature, permittivity, period, and voltage ratio. The M4! command reports the sensor calculated values even if those values are likely to be erroneous.

The M3! command uses several logical tests built into the operating system to ensure that the sensors do not report a number that is likely to be erroneous. Erroneous readings are either outside the sensor operational limits or outside of published accuracy specifications. When one of the following conditions occurs, the logical test replaces the calculated value with another value such as 9999999 or NAN.

Permittivity < 0 or > 88

The equation used to convert period average and electrical conductivity values to permittivity is a three-dimensional surface with two independent variables and eleven coefficients, plus an offset.

Combinations of period and electrical conductivity result in a permittivity calculation that is less than zero or greater than 88. These rare combinations are not expected when the sensor is in soil, but if they do occur, the M3! command reports NAN for permittivity.

Permittivity < 1

If the CS650 calculates a permittivity value greater than zero but less than 1, the M3! command reports a permittivity value of 1.

Permittivity too low for the Topp et al equation

The Topp et al (1980) equation used by the CS650 to estimate volumetric water content works well for most mineral soils. However, at low water contents, the Topp equation may report a negative value for volumetric water content. When the estimated permittivity is greater than 1 but less than 1.881, the M3! command reports a volumetric water content value of zero.

Permittivity more than 42

The Topp et al (1980) equation used by the CS650 to estimate volumetric water content works well for most mineral soils up to a maximum water content of about 0.45. If the CS650 estimates the soil permittivity to be more than 42, which calculates to a volumetric water content of 0.52, then the M3! command reports 9999999 or NAN for volumetric water content.

Calculated permittivity is less than 80% of the permittivity limit

A permittivity limit based on the bulk electrical conductivity (EC) reading is used to determine whether the bulk EC at saturation exceeds the sensor operational limit. That permittivity limit is calculated and compared to the permittivity reading. If the measured permittivity is more than 20% beyond the permittivity limit, the M3! command reports NAN or 9999999 for both permittivity and volumetric water content. This is the most common cause of NAN values with the CS650-series sensors, and it occurs because of soil properties and not because of a sensor malfunction.

Bulk electrical conductivity (EC) is too high

When bulk electrical conductivity is greater than 1.14 dS/m, the solution EC is greater than 3 dS/m, which is the upper limit for accurate readings with the CS650. For the CS655, the upper limit for bulk EC is 3.04 dS/m, corresponding to a solution EC of 8 dS/m. When this bulk EC condition occurs, the soil is considered out-of-bounds and the M3! command reports NAN or 9999999 for both permittivity and volumetric water content.

Voltage ratio is too high

When the voltage ratio is greater than 17, bulk electrical conductivity readings become unstable. When this bulk EC condition occurs, the **M3**! command reports NAN or 9999999 for both permittivity, volumetric water content, and bulk EC.

8.2.2 Use of multiplexers

Multiplexers such as Campbell Scientific AM16/32B can connect up to 32 CS650 sensors to a single control or **U** terminal. When using multiplexers, the simplest configuration is for all sensors to have the same SDI-12 address.

When multiplexing CS650 sensors, use the switched 12V terminal so power to the sensor is turned off under program control before the multiplexer switches to the next terminal.

CAUTION:

Failure to turn off the switched 12 Volt terminal before clocking the multiplexer damages the multiplexer relays.

The proper sequence in the data logger program for measuring CS650 sensors on a multiplexer is:

- 1. Set RES control port high to enable multiplexer
- 2. Pulse CLK control port to advance to next multiplexer terminal
- 3. Set switched 12 volt terminal high to supply power to CS650
- 4. Delay for 3 seconds for sensor to warm up
- 5. Send SDI-12 command(s) to CS650
- 6. Set switched 12 volt terminal low to remove power from CS650
- 7. Repeat steps 2 through 5 for each CS650 connected to the multiplexer
- 8. Set RES control port low to disable multiplexer

Program examples in Example programs (p. 36) show the commands used in CRBasic.

8.3 Water content reflectometer method for measuring volumetric water content

8.3.1 Description of measurement method

For the water content measurement, a differential emitter-coupled logic (ECL) oscillator on the circuit board is connected to the two parallel stainless steel rods. The differentially driven rods form an open-ended transmission line in which the wave propagation velocity is dependent upon the dielectric permittivity of the media surrounding the rods. An ECL oscillator state change is triggered by the return of a reflected signal from the end of one of the rods.

The fundamental principle for CS650 water content measurement is that the velocity of electromagnetic wave propagation along the sensor rods is dependent on the dielectric permittivity of the material surrounding the rods. As water content increases, the propagation velocity decreases because of increasing dielectric permittivity. Therefore, the two-way travel time of the rod signal is dependent upon water content, hence the name water content reflectometer. Digital circuitry scales the high-speed oscillator output to an appropriate frequency for measurement by an onboard microprocessor. Increases in oscillation period resulting from signal attenuation are corrected using an electrical conductivity measurement. A calibration equation converts period and electrical conductivity to bulk dielectric permittivity. The Topp equation is used to convert from permittivity to volumetric water content.

8.3.2 Topp equation

The relationship between dielectric permittivity and volumetric water content in mineral soils has been described by Topp et al. (1980) in an empirical fashion using a third-degree polynomial. With θ_v the volumetric water content and K_a the bulk dielectric permittivity of the soil, the equation presented by Topp et al. is

 $\theta_{\rm v} = -5.3 \cdot 10^{-2} + 2.92 \cdot 10^{-2} {\rm K_a} - 5.5 \cdot 10^{-4} {\rm K_a}^2 + 4.3 \cdot 10^{-6} {\rm K_a}^3$

Research has shown that this equation works well in most mineral soils, so a soil specific calibration of the CS650 sensor is usually not necessary. For a soil specific calibration, you can generate an equation relating K_a to θ_v following the methods described in Water content reflectometer user-calibration (p. 26).

8.3.3 Electrical conductivity

8.3.3.1 Soil electrical conductivity

The quality of soil water measurements which apply electromagnetic fields to wave guides is affected by soil electrical conductivity. The propagation of electromagnetic fields in the configuration of the CS650 is predominantly affected by changing dielectric permittivity due to changing water content, but it is also affected by electrical conductivity. Free ions in soil solution provide electrical conduction paths which result in attenuation of the signal applied to the waveguides. This attenuation both reduces the amplitude of the high-frequency signal on the sensor rods and reduces the bandwidth. The attenuation reduces oscillation frequency at a given water content because it takes a longer time to reach the oscillator trip threshold.

It is important to distinguish between soil bulk electrical conductivity and soil solution electrical conductivity. Soil solution electrical conductivity refers to the conductivity of the solution phase of soil. In the laboratory, extraction methods can determine soil solution electrical conductivity,

 $\sigma_{solution}$. These extraction methods separate the solution from the solid then measures the electrical conductivity of the extracted solution.

The following describes the relationship between solution and bulk electrical conductivity (Rhoades et al., 1976)

 $\sigma_{\text{bulk}} = \sigma_{\text{solution}} \theta_{\text{v}} T + \sigma_{\text{solid}}$

where σ_{bulk} is the electrical conductivity of the bulk soil; σ_{solution} , the soil solution; σ_{solid} , the solid constituents; θ_{v} , the volumetric water content; and T, a soil-specific transmission coefficient intended to account for the tortuosity of the flow path as water content changes. See Rhoades et al., 1989 for a form of this equation which accounts for mobile and immobile water. This publication also discusses soil properties related to CS650 operation such as clay content and compaction. The above equation is presented here to show the relationship between soil solution electrical conductivity and soil bulk electrical conductivity.

Most expressions of soil electrical conductivity are given in terms of solution conductivity or electrical conductivity from extract since it is constant for a soil. Bulk electrical conductivity increases with water content so comparison of the electrical conductivity of different soils must be at the same water content.

The calibration equation in the CS650 operating system corrects the oscillation frequency for the effects of $\sigma_{solution}$ up to 3 dS m⁻¹ for the CS650 and up to 10 dS m⁻¹ for the CS655. This is equivalent to σ_{bulk} values of approximately 0.8 dS m⁻¹ and 2.7 dS m⁻¹ respectively. If σ_{bulk} exceeds these limits, the CS650 sensor returns 99999 for dielectric permittivity and volumetric water content. The measured period average and voltage ratio values continue to report even if the bulk EC is outside the operational range of the sensor.

8.3.3.2 Temperature correction of soil electrical conductivity

The EC value reported by the CS650 is bulk electrical conductivity. This value is temperature dependent, changing by 2% per degree Celsius. To compensate for the effect of temperature, convert EC readings to a standard temperature, such as 25 °C using the following equation:

 $EC_{25} = EC_T / (1 + 0.02 \cdot (T_{soil} - 25))$

where EC₂₅ is the σ_{bulk} value at 25 °C and ECT is the σ_{bulk} value at soil temperature T_{soil} (°C).

8.3.4 Error sources in water content reflectometer measurement

8.3.4.1 Sensor-to-sensor variability error

All manufactured CS650s/CS655s are checked in standard media to develop a sensor specific span and offset value for electrical conductivity and dielectric permittivity measurements. These

sensor specific values are written to the sensor operating system and minimize sensor-to-sensor variability.

8.3.4.2 Insertion error

The method used for sensor insertion can affect the accuracy of the measurement. The sensor rods should be kept as close to parallel as possible when inserted to maintain the design wave guide geometry. The sensitivity of this measurement is greater in the regions closest to the rod surface than at distances away from the surface. Sensors inserted in a manner that generates air voids around the rods indicate lower water content than actual. In some applications, the CS650G insertion and pilot tool helps maintain the proper spacing and parallel orientation of the rods during sensor insertion.

8.3.5 Temperature dependence and correction

The two temperature dependent sources of error in CS650 water content measurements are the effect of temperature on the operation of the sensor electronics and the effect of temperature on the dielectric permittivity of the soil.

The effect of temperature on sensor electronics is minimal with period average readings varying by less than 0.5% of the 20 °C reading over the range of 10 to 30 °C and less than 2% of the 20 °C reading over the range of –10 to 70 °C.

The larger error is caused by the change in dielectric permittivity of soil with temperature. This is mostly due to the high temperature dependence of the permittivity of water, which varies from a value of 88 at 0 °C to 64 at 70 °C. Since water is the major contributor to bulk dielectric permittivity of soil, temperature related changes to the permittivity of water lead to overestimation of volumetric water content at temperatures below 20 °C and underestimation of volumetric water content at temperatures above 20 °C.

The Topp equation does not account for soil temperature. The effect of temperature on the soil permittivity is related to soil specific properties such as porosity and the permittivity of the soil solid phase with temperature. Consequently, a general equation that corrects volumetric water content for temperature for all soils is not available.

A temperature correction equation that works well in quartz sand is given by:

 $\theta_{Corr} = \theta - 0.0044 \bullet T\theta^3 + 0.0014 \bullet T\theta^2 + 0.0029 \bullet T\theta - 0.0002 \bullet T + 2.4 \bullet \theta^3 - 1.6 \bullet \theta^2 + 0.32 \bullet \theta - 0.046$

where θ_{Corr} is the temperature corrected volumetric water content, T is soil temperature in °C, and θ is the volumetric water content value at soil temperature T.

8.3.5.1 Accurate soil temperature measurement

The thermistor used for measuring soil temperature is located in the sensor head and is in contact with one of the stainless steel rods. To make an accurate soil temperature measurement, the sensor head should be buried in the soil so that it is insulated from diurnal temperature fluctuations.

8.4 Water content reflectometer usercalibration

8.4.1 Need for soil specific calibration equation

While the Topp equation works well in a wide range of mineral soils, there are soils for which a user-derived calibration optimizes accuracy of the volumetric water content measurement. The Topp equation underestimates the water content of some organic, volcanic, and fine textured soils. Additionally, porous media with porosity greater than 0.5 or bulk density greater than 1.55 g cm⁻³ may require a media-specific calibration equation.

In these cases, the user may develop a calibration equation to convert CS650 permittivity to volumetric water content over the range of water contents the sensor is expected to measure.

8.4.2 User-derived calibration equation

A quadratic equation or third order polynomial can describe the relationship between soil permittivity and volumetric water content. In many applications, a linear equation similar to Ledieu et al (1986) gives required accuracy.

Quadratic form:

 $\theta_v(K_a) = C_0 + C_1 \cdot K_a + C_2 \cdot K_a^2$

with θ_v the volumetric water content, K_a the bulk dielectric permittivity of the soil, and Cn, the calibration coefficient.

Third-degree polynomial form:

 $\theta_{v}(K_{a}) = C_{0} + C_{1} \bullet K_{a} + C_{2} \bullet K_{a}^{2} + C_{3} \bullet K_{a}^{3}$

with θ_v the volumetric water content, K_a the bulk dielectric permittivity of the soil, and Cn, the calibration coefficient.

Linear form:

 $\theta_v(K_a) = C_0 + C_1 \cdot K_a^{0.5}$

with θ_v the volumetric water content, K_a the bulk dielectric permittivity of the soil, and Cn, the calibration coefficient.

Two data points from careful measurements sometimes are enough to derive a linear calibration. Use at least three data points for a quadratic calibration. With three evenly spaced water contents covering the expected range, the middle water content data point indicates whether a linear or polynomial calibration equation is needed.

Use at least four data points for the derivation of a third-degree polynomial. Space the data points as evenly as practical over the expected range of water content and include the wettest and driest expected values.

8.4.3 Collecting laboratory data for calibration

Water content reflectometer data needed for CS650 calibration are the CS650 permittivity reading and an independently determined volumetric water content. From this data, a linear or polynomial function can describe the sensor response to changing water content. For more information, refer to User-derived calibration equation (p. 26).

Required equipment:

- CS650 connected to data logger programmed to measure permittivity
- Cylindrical sampling devices to determine sample volume for bulk density, such as copper tubing with diameter of ≥ 1 inch and length at about 2 inches.
- Non-metal container such as a PVC pipe (20 cm diameter, 35 cm length) with one end closed. The container should be large enough to ensure that only soil is within 10 cm (4 in) of the CS650 rod surface.
- Oven or microwave safe container of known weight
- Scale to measure soil sample mass
- Oven or microwave to dry samples

The calibration coefficients are derived from a curve fit of known water content and sensor permittivity output. The number of data sets needed to derive a calibration depends on the form of the calibration equation. At least three data sets should be generated to determine whether the linear form is valid. If a polynomial is to be used, four data sets determine whether the function is a quadratic or third order polynomial. Accuracy requirements may need additional data sets. Consider the expected range of soil water content and include data sets from the highest and lowest expected water contents.

Calibration procedure:

- 1. Use relatively dry soil (volumetric water content <10%), which ensures that the soil is homogeneous around the sensor rods.
- 2. Uniformly pack the non-metal container with soil. Compaction of the calibration soil needs to have similar bulk density as the soil at the field site. Follow this method to ensure the soil is compacted uniformly:
 - a. Roughly separate the soil into three or more equal portions.
 - b. Evenly place a loose soil layer in the bottom of the container.
 - c. Tamp the soil to a level that is correct for the target bulk density. Dry soil without compaction has a typical bulk density, 1.1 to 1.4 g cm⁻³. This step is especially important when bulk density is greater than 1.55 g cm⁻³.
 - d. Scarify (loosen) the top of the compacted layer.
 - e. Repeat steps b through d for each of the remaining soil layers.
- 3. Carefully insert CS650 rods through the soil surface until the rods are completely surrounded by soil. Avoid moving the rods from side to side because this can form air voids around the rods.
- 4. Collect the sensor permittivity output and remove sensor.
- 5. Repeat steps 3 and 4 three or four times.
- 6. Add water to top of the container.
- 7. Cover the container to prevent evaporation.
- 8. Frequently observe the CS650 permittivity output, waiting for the permittivity to be constant. This indicates equilibration. The time required for equilibration depends on the amount of water added and the hydraulic properties of the soil.
- 9. Record the CS650 permittivity.
- 10. Repeat steps 6 through 9.
- 11. Remove the CS650.
- 12. Take core samples of soil using this process:
 - a. Evenly push copper tubing into the soil where the sensor rods were inserted.
 - b. Remove the tube containing the sample and gently trim the ends of excess soil.
 - c. Remove excess soil from outside of the tube.
 - d. Transfer the soil from the tube to the oven or microwave safe container.

- e. Weigh the container holding the wet soil, and subtract the container weight from this value to get the wet soil mass.
- f. Record the wet soil mass.
- g. Use oven or microwave to remove water from soil. Oven drying requires 24 hours at 105 °C. Microwave drying typically takes 20 minutes depending on microwave power and sample water content. ASTM Method D4643-93 requires heating in microwave for 3 minutes, cooling in desiccator then weighing and repeating this process until measured mass is constant (References (p. 34)).
- h. Weigh the container holding the dry soil, and subtract the container weight from this value to get the dry soil mass.
- i. Record the dry soil mass.
- j. Repeat process to take replicate samples. Three carefully handled samples provide good results.
- 13. Calculate gravimetric water content from wet mass of sample (m_{wet}) and dry mass of sample (m_{drv}):

$$heta_{
m g} = rac{{
m m}_{
m wet} - {
m m}_{
m dry}}{{
m m}_{
m dry}}$$

For the bulk density

$$ho_{
m bulk} = rac{
m m_{dry}}{
m volume_{cylinder}}$$

the dry mass of the sample is divided by the sample tube volume.

The volumetric water content is the product of the gravimetric water content and the bulk density

$$heta_{\mathrm{v}} = heta_{\mathrm{g}} ullet
ho_{\mathrm{bulk}}$$

The average water content for the replicates and the recorded CS650 permittivity are one datum pair to be used for the calibration curve fit.

8.4.4 Collecting field data for calibration

Required equipment:

- CS650 connected to data logger programmed to measure sensor permittivity
- Cylindrical sampling devices to determine sample volume for bulk density, such as copper tubing with diameter of ≥ 1 inch and length about 2 inches.
- Oven or microwave safe container of known weight

- Scale to measure soil sample mass
- Oven or microwave to dry samples

Data needed for CS650 calibration are the CS650 permittivity output and an independently determined volumetric water content. From this data, the sensor response to changing water content can be described by a function as described in User-derived calibration equation (p. 26).

The calibration coefficients are derived from a curve fit of known water content and sensor permittivity output. The number of data sets needed to derive a calibration depends on the form of the calibration equation. At least three data sets should be generated to determine whether the linear form is valid. If a polynomial is to be used, four data sets determine whether the function is a quadratic or third order polynomial. Accuracy requirements may need additional data sets. Consider the expected range of soil water content and include data sets from the highest and lowest expected water contents.

Collecting measurements of CS650 permittivity and core samples from the location where the sensor is used provides the best on-site soil-specific calibration. However, intentionally changing water content in soil profiles is difficult.

Soil hydraulic properties are spatially variable. Obtaining measurements that are representative of the soil on a large scale requires multiple readings and sampling. The average of several core samples should be used to calculate volumetric water content. Likewise, the CS650 should be inserted at least three times into the soil recording the permittivities following each insertion and using the average.

The following is the procedure for obtaining samples.

- 1. Use a shovel to form a vertical face of soil.
- 2. If using the CS650 within about 0.5 metres of the surface, insert the sensor into the face.
- 3. Add water to the surface using percolation.
- 4. Frequently observe the CS650 permittivity output, waiting for the permittivity to be constant. This indicates equilibration.
- 5. Record the CS650 permittivity.
- 6. Repeat steps 2 through 5 at least three times.
- 7. Remove the CS650.
- 8. Take core samples of the soil using this process:
 - a. Evenly push copper tubing into the soil surface.
 - b. Remove the tube containing the sample and gently trim the ends of excess soil.

- c. Remove excess soil from outside of tube.
- d. Transfer the soil from the tube to the oven or microwave safe container.

NOTE:

If samples must be stored prior to weighing, seal the container with tape or inside a plastic bag to prevent water loss and store away from direct sunlight.

- e. Weigh the container holding the wet soil, and subtract the container weight from this value to get the wet soil mass.
- f. Record the wet soil mass.
- g. Remove water from the sample by heating using an oven or microwave. Oven drying requires 24 hours at 105 °C. Microwave drying typically takes 20 minutes depending on microwave power and sample water content. ASTM Method D4643-93 requires heating in microwave for 3 minutes, cooling in desiccator then weighing and repeating this process until mass is constant.
- h. Weigh the container holding the dry soil, and subtract the container weight from this value to get the dry soil mass.
- i. Record the dry soil mass.
- j. Repeat process to take replicate samples. Three carefully handled samples provide good results.
- 9. Calculate gravimetric water content from wet mass of sample (m_{wet}) and dry mass of sample (m_{drv}):

$$heta_{
m g} = rac{{
m m_{wet}} - {
m m_{dry}}}{{
m m_{dry}}}$$

For the bulk density,

$$ho_{
m bulk} = rac{{
m m}_{
m dry}}{{
m volume}_{
m cylinder}}$$

the dry mass of the sample is divided by the sample tube volume.

The volumetric water content is the product of the gravimetric water content and the bulk density

$heta_{\mathrm{v}} = heta_{\mathrm{g}} ullet ho_{\mathrm{bulk}}$

The average water content for the replicates and the recorded CS650 period are one datum pair to be used for the calibration curve fit.

8.4.5 Calculations

The empty cylinders used for core sampling should be clean and both empty mass and volume are measured and recorded. For a cylinder, the volume is

$$ext{volume} = \pi ullet \left(rac{ ext{d}}{2}
ight)^2 ullet ext{h}$$

where d is the inside diameter of the cylinder and h is the height of the cylinder.

During soil sampling it is important that the cores be completely filled with soil but not extend beyond the ends of the cylinder.

Once soil core samples are obtained, place the soil-filled cylinder in a small tray of known empty mass. This tray holds the core sample during drying in an oven.

To obtain m_{wet} , subtract the cylinder empty mass and the container empty mass from the mass of the soil filled cylinder in the tray. Remove all the soil from the cylinder and place this soil in the tray. Dry the samples using oven or microwave methods as described earlier.

To obtain m_{dry} , weigh the tray containing the soil after drying. Subtract tray mass for m_{dry} . Calculate gravimetric water content, θ_{q} , using

$$heta_{
m g} = rac{{
m m}_{
m wet} - {
m m}_{
m dry}}{{
m m}_{
m dry}}$$

To obtain soil bulk density, use

$$ho_{
m bulk} = rac{
m m_{dry}}{
m volume_{cylinder}}$$

Volumetric water content is calculated using

$$\theta_{\rm v} = \theta_{\rm g} \bullet \rho_{\rm bulk}$$

9. Maintenance and troubleshooting

The CS650 does not require periodic maintenance. Table 9-1 (p. 33) provides troubleshooting information.

Table 9-1: Symptom, cause, and	solutions	
Symptom	Possible cause	Solution
All CS650 output values read 0	No SDI12Recorder instruction in data logger program	Add SDI12Recorder instruction to data logger program
	Conditional statement that triggers reading is not evaluating as true	Check logic of conditional statement that triggers readings
All values read NAN. First value reads NAN and all other values read 0 or never change from one measurement to another	CS650 SDI-12 address does not match address specified in data logger program	Change sensor address or modify program so that they match
	CS650 green wire not attached to SDI port specified in data logger program	Connect wire to correct control port or modify program to match wiring
	CS650 not being powered	Make sure red wire is connected to 12V or SW12V and black wire to G
		If using SW12 to power sensor, make sure red wire is connected and data logger program switches SW12 on
VWC reading is 9999999	Soil bulk permittivity is outside sensor operational range	Modify program to collect permittivity value and try soil specific calibration
EC reading is 9999999	Soil bulk electrical conductivity is outside sensor operational range	If using CS650, try CS655
Readings erratic, including NAN and 9999999	Multiple sensors with same SDI-12 address sharing same control port	Give sensors unique addresses or put on separate control ports

10. References

Ledieu, J., P. De Ridder, P. De Clerck, and S. Dautrebande. 1986. "A method of measuring soil moisture by time-domain reflectometry," J. Hydrol. 88:319-328.

Rhoades, J.D., P.A.C. Raats, and R.J. Prather. 1976. *Effects of liquid-phase electrical conductivity, water content and surface conductivity on bulk soil electrical conductivity*. Soil Sci. Soc. Am. J., 40: 651-653.

Rhoades, J.D., N.A. Manteghi, P.J. Shouse, W.J. Alves. 1989. *Soil electrical conductivity and soil salinity: New formulations and calibrations*. Soil Sci. Soc. Am. J., 53:433-439.

Topp, G.C., J.L. Davis & A.P. Annan. 1980. "*Electromagnetic determination of soil water content: measurements in coaxial transmission lines*," Water Resources Research, v. 16, No. 3:574-582.

Appendix A. Importing *Short Cut* code into *CRBasic Editor*

Short Cut creates a .DEF file that contains wiring information and a program file that can be imported into the *CRBasic Editor*. By default, these files reside in the C:\campbellsci\SCWin folder.

Import *Short Cut* program file and wiring information into *CRBasic Editor*.

 Create the *Short Cut* program. After saving the *Short Cut* program, click the *Advanced* tab then the *CRBasic Editor* button. A program file with a generic name will open in CRBasic. Provide a meaningful name and save the CRBasic program. This program can now be edited for additional refinement.

NOTE:

Once the file is edited with *CRBasic Editor*, *Short Cut* can no longer be used to edit the program it created.

- 2. To add the *Short Cut* wiring information into the new CRBasic program, open the .DEF file located in the C:\campbellsci\SCWin folder, and copy the wiring information, which is at the beginning of the .DEF file.
- 3. Go into the CRBasic program and paste the wiring information into it.
- 4. In the CRBasic program, highlight the wiring information, right-click, and select **Comment Block**. This adds an apostrophe (') to the beginning of each of the highlighted lines, which instructs the data logger compiler to ignore those lines when compiling. The **Comment Block** feature is demonstrated at about 5:10 in the CRBasic | Features video .

Appendix B. Example programs

B.1 CR1000X with one CS650 sensor

This CRBasic example program measures one CS650 sensor on a CR1000X every 15 minutes, storing hourly averages of volumetric water content, electrical conductivity, and soil temperature and samples of permittivity, period average and voltage ratio. The CS650 has an SDI-12 address of 0. Wiring for the example is shown in Table B-1 (p. 36).

Table B-1: CR1000X wiring for one sensor example program		
CR1000X CS650		
12V	Red	
C1	Green	
G	Black, orange, clear	

CRBasic Example 1: CR1000X with one CS650 sensor

```
Public CS650(6)
'Assign aliases to the public array
Alias CS650(1)=VWC: Alias CS650(2)=EC: Alias CS650(3)=TSoil
Alias CS650(4)=Perm: Alias CS650(5)=PerAvg: Alias CS650(6)=VoltR
Units VWC = m^3/m^3: Units EC = dS/m: Units TSoil = deg C
DataTable (DatoutCS650,1,-1)
 DataInterval (0,60,Min,10)
 Average (3,CS650(1),FP2,False)
 Sample (3,CS650(4),IEEE4)
EndTable
BeginProg
 Scan (15, Min, 0, 0)
    SDI12Recorder (CS650(),C1,0,"M4!",1.0,0)
   CallTable DatoutCS650 'Call Data Table
 NextScan
EndProg
```

B.2 CR1000X with two CS650 sensors on same control port

This CRBasic example program measures two CS650 sensors on a CR1000X every 15 minutes, storing hourly averages of volumetric water content, electrical conductivity, and soil temperature and samples of permittivity, period average and voltage ratio. The first CS650 has an SDI-12 address of 0 and the second has an address of 1. Wiring for the example is shown in Table B-2 (p. 37). Assignment of aliases and units is not shown in this example.

Table B-2: CR1000X wiring for two sensor example program		
CR1000X	CS650 (wiring same for both)	
12V	Red	
C1	Green	
G	Black, orange, clear	

CRBasic Example 2: CR1000X with two CS650 sensors on same control port
Public CS650(6) Public CS650_2(6) DataTable (DatoutCS650,1,-1) DataInterval (0,60,Min,10) Average (3,CS650(1),FP2,False) Sample (3,CS650(4),IEEE4) Average (3,CS650_2(1),FP2,False) Sample (3,CS650_2(4),IEEE4)
EndTable
Scan (15,Min,0,0)
SDI12Recorder (CS650(),C1,0,"M4!",1.0,0) 'Measure CS650 with address 0 SDI12Recorder (CS650_2(1),C1,1,"M4!",1.0,0) 'Measure CS650 with address 1 CallTable DatoutCS650 'Call Data Table NextScan EndProg

B.3 CR1000X with 12 CS650 sensors on multiplexer

This CRBasic example program measures 12 CS650 sensors on a AM16/32B multiplexer every 15 minutes, storing hourly averages of volumetric water content, electrical conductivity, soil temperature, permittivity, period average, and voltage ratio. All sensors are addressed with SDI-12 address of 0. In this example, the sensors are powered through the switched 12V terminal and require 3 seconds warm-up time per sensor. Total time to measure all 12 sensors is more than 36 seconds. Alternately, all of the red wires for the sensors could be connected to a bus separate from the multiplexer with the bus connected to 12V for continuous power. This would decrease measurement time. Wiring for the example is shown in TABLE B 3. Assignment of aliases and units is not shown in this example.

Table B-3: CR1000X wiring for multiplexer example program			
CR1000X	AM16/32B (2x32 mode)	CS650	
12V	12V		
G	GND		
C2	RES		
C3	CLK		
SW12	COM ODD H		
C1	COM ODD L		
G	COM Ground		
	High terminals 1H – 12H	Red	
	Low terminals 1L – 12L	Green	
	Ground terminals to left of low terminals	Black, orange, clear	

CRBasic Example 3: CR1000X with 12 CS650 sensors on multiplexer

```
Dim LCount
Public CS650(12,6)
DataTable (DatoutCS650,1,-1)
 DataInterval (0,60,Min,2)
 Average (72,CS650(),IEEE4,False)
EndTable
BeginProg
 Scan (15,Min,0,0)
   PortSet(C2,1) 'Turn AM16/32 Multiplexer On
   Delay(0, 150, mSec)
   LCount=1
   SubScan(0,uSec,12)
      PulsePort(C3,10000 'Switch to next AM16/32 terminal
      SW12 (SW12_1,1) 'Apply power to CS650
      Delay (0,3,Sec) 'Wait three seconds for sensor to warm up
      SDI12Recorder (CS650(LCount,1),C1,0,"M4!",1.0,0)
      LCount=LCount+1
      SW12 (SW12_1,0) 'Remove power from CS650
   NextSubScan
   PortSet(C2,0) 'Turn AM16/32 Multiplexer Off
   Delay(0, 150, mSec)
   CallTable DatoutCS650 'Call Data Table
 NextScan
EndProg
```

Appendix C. Discussion of soil water content

The water content reflectometer measures volumetric water content. Soil water content is expressed on a gravimetric and a volumetric basis. To obtain the independently determined volumetric water content, gravimetric water content must first be measured. Gravimetric water content (θ g) is the mass of water per mass of dry soil. It is measured by weighing a soil sample (m_{wet}), drying the sample to remove the water, then weighing the dried soil (m_{dry}).

$$heta_{
m g} = rac{{
m m_{water}}}{{
m m_{soil}}} = rac{{
m m_{wet}} - {
m m_{dry}}}{{
m m_{dry}}}$$

Volumetric water content (θ_v) is the volume of liquid water per volume of soil. Volume is the ratio of mass to density (ρ_b) which gives:

$$heta_{\mathrm{v}} = rac{\mathrm{volume_{water}}}{\mathrm{volume_{soil}}} = rac{rac{\mathrm{m_{water}}}{
ho_{\mathrm{water}}}}{rac{\mathrm{m_{soil}}}{
ho_{\mathrm{soil}}}} = rac{ heta_{\mathrm{g}} ullet
ho_{\mathrm{soil}}}{
ho_{\mathrm{water}}}$$

The density of water is close to 1 and often ignored.

Soil bulk density (ρ_{bulk}) is used for ρ_{soil} and is the ratio of soil dry mass to sample volume.

$$ho_{
m bulk} = rac{{
m m}_{
m dry}}{{
m volume}_{
m sample}}$$

Another useful property, soil porosity (ϵ), is related to soil bulk density as shown by the following expression.

$$arepsilon = 1 - rac{
ho_{ ext{bulk}}}{
ho_{ ext{solid}}}$$

The term ρ_{solid} is the density of the soil solid fraction and is approximately 2.65 g cm⁻³.

Appendix D. SDI-12 sensor support

SDI-12, Serial Data Interface at 1200 baud, is a protocol developed to simplify sensor and data logger compatibility. Only three wires are necessary — serial data, ground, and 12 V. With unique addresses, multiple SDI-12 sensors can connect to a single SDI-12 terminal on a Campbell Scientific data logger.

This appendix discusses the structure of SDI-12 commands and the process of querying SDI-12 sensors. For more detailed information, refer to version 1.4 of the SDI-12 protocol, available at www.sdi-12.org

For additional information, refer to the SDI-12 Sensors | Transparent Mode and SDI-12 Sensors | Watch or Sniffer Mode videos.

D.1 SDI-12 command basics

SDI-12 commands have three components:

- Sensor address (a) a single character and the first character of the command. Use the default address of zero (0) unless multiple sensors are connected to the same port.
- Command body an upper case letter (the "command"), optionally followed by one or more alphanumeric qualifiers.
- Command termination (!) an exclamation mark.

An active sensor responds to each command. Responses have several standard forms and always terminate with <CR><LF> (carriage return and line feed). Standard SDI-12 commands are listed in Table D-1 (p. 41).

Table D-1: Campbell Scientific sensor SDI-12 command and response set			
Name	Command	Response ¹	
Acknowledge Active	a!	a <cr><lf></lf></cr>	
Send Identification	aI!	allcccccccmmmmmvvvxxxxx <cr><lf></lf></cr>	
Start Verification	aV!	atttn <cr><lf></lf></cr>	

Table D-1: Campbell Scientific sensor SDI-12 command and response set			
Name	Command	Response ¹	
Address Query	?!	a <cr><lf></lf></cr>	
Change Address	aAb!	b <cr><lf></lf></cr>	
Start Measurement	aM! aM1!aM9!	atttn <cr><lf></lf></cr>	
Send DataaD0!aD9!a <values><cr><lf> or a<values><crc><cr><lf></lf></cr></crc></values></lf></cr></values>			
¹ Information on each of these commands is given in the following sections			

D.1.1 Acknowledge active command (a!)

The Acknowledge Active command (a!) is used to test a sensor on the SDI-12 bus. An active sensor responds with its address.

D.1.2 Send identification command (al!)

Sensor identifiers are requested by issuing command **aI!**. The reply is defined by the sensor manufacturer but usually includes the sensor address, SDI-12 version, manufacturer's name, and sensor model information. Serial number or other sensor specific information may also be included.

aI!	allcccccccmmmmmvvvxxxxx <cr><lf></lf></cr>
а	Sensor SDI-12 address
II	SDI-12 version number (indicates compatibility)
сссссссс	8-character vendor identification
mmmmmm	6 characters specifying the sensor model
VVV	3 characters specifying the sensor version (operating system)
xxxxx	Up to 13 optional characters used for a serial number or other specific sensor information that is not relevant for operation of the data logger
<cr><lf></lf></cr>	Terminates the response
Source: SDI-12: A Serial-D	igital Interface Standard for Microprocessor-Based Sensors (see References).

D.1.3 Start verification command (aV!)

The response to a Start Verification command can include hardware diagnostics, but like the **aI**! command, the response is not standardized.

Command: aV!

Response: atttn<CR><LF>

a = sensor address

ttt = time, in seconds, until verification information is available

n = the number of values to be returned when one or more subsequent **D**! commands are issued

D.1.4 Address query command (?!)

Command **?!** requests the address of the connected sensor. The sensor replies to the query with the address, *a*. This command should only be used with one sensor on the SDI-12 bus at a time.

D.1.5 Change address command (aAb!)

Multiple SDI-12 sensors can connect to a single SDI-12 terminal on a data logger. Each device on a single terminal must have a unique address.

A sensor address is changed with command aAb!, where *a* is the current address and *b* is the new address. For example, to change an address from 0 to 2, the command is 0A2!. The sensor responds with the new address *b*, which in this case is 2.

NOTE:

Only one sensor should be connected to a particular terminal at a time when changing addresses.

D.1.6 Start measurement commands (aM!)

A measurement is initiated with the M! command. The response to each command has the form atttn < CR > < LF >, where

a = sensor address

ttt = time, in seconds, until measurement data is available. When the data is ready, the sensor notifies the data logger, and the data logger begins issuing **D** commands.

n = the number of values returned when one or more subsequent **D** commands are issued. For the **aM!** command, n is an integer from 0 to 9.

When the **aM!** is issued, the data logger pauses its operation and waits until either it receives the data from the sensor or the time, *ttt*, expires. Depending on the scan interval of the data logger program and the response time of the sensor, this may cause skipped scans to occur. In this case make sure your scan interval is longer than the longest measurement time (*ttt*).

Table D-2: Example aM! s	equence
OM !	The data logger makes a request to sensor 0 to start a measurement.
00352 <cr><lf></lf></cr>	Sensor 0 immediately indicates that it will return two values within the next 35 seconds.
0 <cr><lf></lf></cr>	Within 35 seconds, sensor 0 indicates that it has completed the measurement by sending a service request to the data logger.
0D0 !	The data logger immediately issues the first D command to collect data from the sensor.
0+.859+3.54 <cr><lf></lf></cr>	The sensor immediately responds with the sensor address and the two values.

D.1.7 Stopping a measurement command

A measurement command (M!) is stopped if it detects a break signal before the measurement is complete. A break signal is sent by the data logger before most commands.

A concurrent measurement command (C!) is aborted when another valid command is sent to the sensor before the measurement time has elapsed.

D.1.8 Send data command (aD0! ... aD9!)

The Send Data command requests data from the sensor. It is issued automatically with every type of measurement command (aM!, aMC!, aC!, aC!). When the measurement command is aM! or aMC!, the data logger issues the aDO! command once a service request has been received from the sensor or the reported time has expired. When the data logger is issuing concurrent commands (aC! or aCC!), the Send Data command is issued after the required time has elapsed (no service request will be sent by the sensor). In transparent mode (see SDI-12 transparent mode (p. 45)), the user asserts this command to obtain data.

Depending on the type of data returned and the number of values a sensor returns, the data logger may need to issue **aD0**! up to **aD9**! to retrieve all data. A sensor may return up to 35 characters of data in response to a **D** command that follows an **M**! or **MC**! command. A sensor may return up to 75 characters of data in response to a **D** command that follows a **C**! or **CC**! command. Data values are separated by plus or minus signs.

Command: aD0! (aD1! ... aD9!)

Response: *a*<*values*><*CR*><*LF*> or *a*<*values*><*CRC*><*CR*><*LF*>

where:

a = the sensor address

<values> = values returned with a polarity sign (+ or –)

<*CR*><*LF*> = terminates the response

< CRC > = 16-bit CRC code appended if data was requested with aMC! or aCC!.

D.2 SDI-12 transparent mode

System operators can manually interrogate and enter settings in probes using transparent mode. Transparent mode is useful in troubleshooting SDI-12 systems because it allows direct communication with probes. Data logger security may need to be unlocked before activating the transparent mode.

Transparent mode is entered while the computer is communicating with the data logger through a terminal emulator program. It is accessed through Campbell Scientific data logger support software or other terminal emulator programs. Data logger keyboards and displays cannot be used.

The terminal emulator is accessed by navigating to the **Tools** list in *PC400* or the **Datalogger** list in the **Connect** screen of *LoggerNet*.

Watch videos/sdi12-sensors-transparent-mode 🕨 from our website.

Data loggers from other manufacturers will also have a transparent mode. Refer to those manuals on how to use their transparent mode.

The following examples show how to enter transparent mode and change the SDI-12 address of an SDI-12 sensor. The steps shown in Changing an SDI-12 address (p. 45) are used with most Campbell Scientific data loggers.

D.2.1 Changing an SDI-12 address

This example was done with a CR1000X, but the steps are only slightly different for CR6, CR3000, CR800-series, CR300-series, CR1000 data loggers.

- 1. Connect an SDI-12 sensor to the CR1000X.
- 2. Open *Device Configuration* utility.
- 3. Under **Device Type**, type the data logger model and double-click on the model type. This example uses a CR1000X directly connected to the computer USB port.

4. Select the correct Communication Port and click Connect.



5. Click the **Terminal** tab.

O Device Configuration Utility 2.22									-	×
File Backup Options Help No	w Version ★									
Device Type	Deployment	Logger Control	Data Monitor	Data Collection	File Control	Manage OS ★	Settings Editor	Terminal		
Q Search										
CR 1000										
CR 1000X Series										
CR300 Series										
CR3000										
CR6 Series										

6. Select All Caps Mode.

All Caps	Echo Input	Pause	Start Export	Send File	

7. Press Enter until the data logger responds with the data logger (CR1000X>) prompt.

Device Configuration Utilit File Backup Options Help	ty 2.22 p New Version ★	-	×
Device Type	Deployment Logger Control Data Monitor Data Collection File Control Manage OS * Settings Editor Termi	ıal	
Q Search			
CR 1000	^ CRIOOOX>		
CR 1000X Series			
CR300 Series			
CR3000			
CR6 Series			

- 8. Type SDI12 and press Enter.
- 9. At the **Select SDI12 Port** prompt, type the number corresponding to the control port where the sensor is connected and press Enter. In this example the sensor is connected to C3. The

response **Entering SDI12 Terminal** indicates that the sensor is ready to accept SDI-12 commands.

```
CR1000X>
CR1000X>SDI12
1: C1
2: C3
3: C5
4: C7
Select SDI12 Port: 2
```

10. To query the sensor for its current SDI-12 address, type **?!** and press Enter. The sensor responds with its SDI-12 address. If no characters are typed within 60 seconds, the mode is exited. In that case, simply type **SDI12** again, press Enter, and type the correct control port number when prompted.

?! 0

- To change the SDI-12 address, type **aAb!**, where **a** is the current address from the previous step and **b** is the new address. Press **Enter**. The sensor changes its address and responds with the new address. In the following example, the sensor address is changed from 0 to B.
 SDI12
 SDI12
 SDI12
- 12. To exit SDI-12 transparent mode, click Close Terminal.

NOTE:

The transparent mode for the CR6, CR3000, CR800-series, CR300-series, and CR1000 data loggers is similar to that shown for the CR1000X.

D.3 References

SDI-12 Support Group. SDI-12: A Serial-Digital Interface Standard for Microprocessor-Based Sensors – Version 1.4. River Heights, UT: SDI-12 Support Group, 2017. https://sdi-12.org/specification 2.



Campbell Scientific Regional Offices

Australia

Location:	Garbutt, QLD Australia
Phone:	61.7.4401.7700
Email:	info@campbellsci.com.au
Website:	www.campbellsci.com.au

Brazil

Location:São Paulo, SP BrazilPhone:11.3732.3399Email:vendas@campbellsci.com.brWebsite:www.campbellsci.com.br

Canada

Location:	Edmonton, AB Canada
Phone:	780.454.2505
Email:	dataloggers@campbellsci.ca
Website:	www.campbellsci.ca

China

Location:	Beijing, P. R. China
Phone:	86.10.6561.0080
Email:	info@campbellsci.com.cn
Website:	www.campbellsci.com.cn

Costa Rica

Location:	San Pedro, Costa Rica
Phone:	506.2280.1564
Email:	info@campbellsci.cc
Website:	www.campbellsci.cc

France

Location:	Vincennes, France
Phone:	0033.0.1.56.45.15.20
Email:	info@campbellsci.fr
Website:	www.campbellsci.fr

Germany

Location:Bremen, GermanyPhone:49.0.421.460974.0Email:info@campbellsci.deWebsite:www.campbellsci.de

India

Location:	New Delhi, DL India
Phone:	91.11.46500481.482
Email:	info@campbellsci.in
Website:	www.campbellsci.in

South Africa

Location:	Stellenbosch, South Africa
Phone:	27.21.8809960
Email:	sales@campbellsci.co.za
Website:	www.campbellsci.co.za

Spain

Location:	Barcelona, Spain
Phone:	34.93.2323938
Email:	info@campbellsci.es
Website:	www.campbellsci.es

Thailand

Location:	Bangkok, Thailand
Phone:	66.2.719.3399
Email:	info@campbellsci.asia
Website:	www.campbellsci.asia

UK

Location:	Shepshed, Loughborough, UK
Phone:	44.0.1509.601141
Email:	sales@campbellsci.co.uk
Website:	www.campbellsci.co.uk

USA

Location:	Logan, UT USA
Phone:	435.227.9120
Email:	info@campbellsci.com
Website:	www.campbellsci.com