



EMS Brno

Data Acquisition Environment

Hardware – Software – Cloud application

[www.emsbrno.cz](http://www.emsbrno.cz)

## Global radiation sensor EMS 11S

### Main features:

- Glass diffuser – long time period between recalibrations
- SDI-12 output
- Silicon semiconductor photovoltaic detector
- Connector in sensor body



### Specification:

- SDI-12 version 1.3 compatible
- Calibration error under the daylight condition max. 7%
- Recommended recalibration each 10 years
- Linearity better than 1%
- Cosine error <10% up to 85° angle of incidence
- Operating Temperature -20 to 60 °C
- Size: diameter 22 mm, length 66 mm




EMS 11S in AL0171

## Specification:

Output	SDI-12
Sensitivity	0.05 mW m <sup>-2</sup>
Cosine error	Less than 10% up to 85° angle of incidence
Linearity	Better than 1%
Connection	Three wire Escha M8 connector female
Size (diameter x length)	22 x 66 mm
Weight	48 g
Operating environment temperature	-40 to 80 °C
Operating environment humidity	0 to 100%
Protection	IP 68

### Global radiation sensor EMS 11S - female connector wiring




### Standard M8 male connector cable wiring



Brown - 6.5-16V  
Black - Data  
Blue - GND

### EMS 11S – cosine characteristic:



## SDI-12 interface specification EMS11S.

Sensor EMS11S is compatible with SDI-12 version 1.3 (more info at <http://sdi-12.org/archives.php>), except for continuous measurements (aR0 - aR9 or aRC0 - aRC9).

### Overview of supported commands:

Datalogger commands are in bold. Each response from the sensor is terminated by <CR> <LF>

#### Info command - **aI!**

For instance:

**1I!** 113EMSBrho EMS11S1.5Sn#1234567890

Parameter	Length	Description
1I!	3	Request to read the sensor information at address 1
1	1	Sensor address - here 1
13	2	SDI version - here 1.3
EMSBrho	8	Manufacturer - completed with space 0x20
EMS11S	6	Model - completed with space 0x20
1.5	3	FW Sensor version - here 1.5
Sn#1234567890	13	Serial number of the sensor

#### Measurement command - **aM!**

For instance:

**1M!** 10011

Parameter	Length	Description
1M!	3	Sensor measurement request at address 1
1	1	Sensor address - here 1
001	3	Time after which the measured data will be available in seconds - here 1. If the data is available earlier, the sensor sends the address terminated by the <CR> <LF> - service request.
1	1	Number of variables returned - here 1

### Data command - aD0!

For instance (1M):

**1D0!** 1+525.32XYZ

Parameter	Length	Description
1D0!	4	Sensor data request at address 1
1	1	Sensor address - here 1
+525.32	Variable	Global radiation in W/m2
XYZ	3	16-bit CRC - added only if aMC! or aCC! commands were requested for the measurement

Note: The sensor returns 9999.9 when the global radiation value is higher than 1500.

### Change Address - aAb!

For instance:

**1A2!** 2

Parameter	Length	Description
1A2!	4	Request to change the sensor address on the address 1 to address 2
2	1	New sensor address - here 2

**Address Query command -?! - Be careful - there must be only one sensor on the line!**

For instance:

**?!** 2

Parameter	Length	Description
?!	2	Retrieving the sensor address
2	1	Attached sensor address - here 2.

### Concurrent Measurement - aC!

For instance:

**1C!** 100101

Parameter	Length	Description
1C!	3	Sensor measurement request at address 1
1	1	Sensor address - here 1
001	3	Time after which the measured data will be available in seconds - here 1
01	2	Number of variables returned - here 1

### Measurement command with CRC - aMC!

For instance:

**1MC!** 10011

Parameter	Length	Description
1MC!	4	Sensor measurement request at address 1 with CRC data control
1	1	Sensor address here 1
001	3	Time after which the measured data will be available in seconds - here 1. If the data is available earlier, the sensor sends the address terminated by the <CR> <LF> - service request.
1	1	Number of variables returned - here 2

### Concurrent Measurement with CRC - aCC!

For instance:

**1CC!** 100101

Parameter	Length	Description
1CC!	4	Sensor measurement request at address 1 with CRC data control
1	1	Sensor address here 1
001	3	Time after which the measured data will be available in seconds - here 1
01	2	Number of variables returned - here 2

### Verification command - aV!

For instance:

**1V!** 10011

Parameter	Length	Description
1V!	3	Sensor measurement request at address 1
1	1	Sensor address here 1
001	3	Time after which the measured data will be available in seconds - here 1
1	1	Number of variables returned - here 2

## Acknowledge Active – a!

For instance:

**2! 2**

Parameter	Length	Description
2!	2	Check the sensor connection
2	1	Sensor address respond - here 2